

**March 18, 2015**

**Subject: Li-Ion Batteries in Thermo Scientific Model TruNarc Series Instruments**

To Whom It May Concern:

The Model TruNarc series instruments manufactured by Thermo Fisher Scientific use Model PT01550 battery packs manufactured by ProTechnologies, Inc. of Pilot Mountain, NC. The battery packs are made with NCR18650 cells manufactured by Panasonic. For this battery, Panasonic provides the attached Product Safety information. ProTechnologies, Inc provides certification of successful completion of the required UN tests for Li-ion battery transport.

If you have any further questions about these batteries, please contact the following:

- ProTechnologies, Inc at 336-368-1375
- Panasonic North and Latin America at 1-800-211-PANA (7262)
- Thermo Fisher Scientific Customer Support at 978-670-7460.

Sincerely,



Jim Blute  
Environmental Health and Safety Manager  
Thermo Scientific Portable Analytical Instruments

Thermo Fisher Scientific  
2 Radcliff Road  
Tewksbury, MA 01876 USA  
Phone: +1 978-215-1310 Mobile: +1 978-790-8269  
jim.blute@thermofisher.com  
[www.thermoscientific.com/pai](http://www.thermoscientific.com/pai)

# Product Information Sheet

## Panasonic Batteries

Panasonic Industrial Company  
 A Division of Panasonic Corporation of North America  
 5201 Tollview Drive, 1F-3  
 Rolling Meadows, IL 60008  
 Toll Free: 877-726-2228  
 Fax: 847-468-5750  
 Internet: [www.panasonic.com/industrial/batteries-oe](http://www.panasonic.com/industrial/batteries-oe)  
 e-mail: [oebatteries@us.panasonic.com](mailto:oebatteries@us.panasonic.com)

**Product:**        **Lithium-ion Batteries  
(Li-ion)**

**Applicable models/sizes:** **All Cylindrical  
and Prismatic Lithium-ion batteries**

**Revision:** – **January 1, 2015**

**The batteries referenced herein are exempt articles and are not subject to the OSHA Hazard Communication Standard requirement. This sheet is provided as a service to our customers.**

### MSDS

Material Safety Data Sheets (MSDS) are a sub-requirement of the Occupational Safety and Health Administration (OSHA) Hazard Communication Standard, 29 CFR Subpart 1910.1200. This Hazard Communication Standard does not apply to various subcategories including anything defined by OSHA as an "article". OSHA has defined "article" as a manufactured item other than a fluid or particle; (i) which is formed to a specific shape or design during manufacture; (ii) which has end use function(s) dependent in whole or in part upon its shape or design during end use; and (iii) which under normal conditions of use does not release more than very small quantities, e.g. minute or trace amounts of a hazardous chemical, and does not pose a physical hazard or health risk to employees.

*Because all of our batteries are defined as "articles", they are exempt from the requirements of the Hazard Communication Standard, hence a MSDS is not required.*

The following components are found in a Panasonic Lithium Ion battery:

#### Nickel Manganese Cobalt Type

Component	Material	Formula / CAS
Positive Electrode	Lithium Nickel Manganese Cobalt Oxide	LiMnCoO <sub>2</sub> 346417-97-8
Negative Electrode	Graphite	C                7440-44-0
Electrolyte	Ethylene Carbonate – Solvent	C <sub>3</sub> H <sub>4</sub> O <sub>3</sub> 96-49-1
	Diethyl Carbonate – Solvent	C <sub>5</sub> H <sub>10</sub> O <sub>3</sub> 105-58-8
	Lithium Hexafluorophosphate – Salt	LiPF <sub>6</sub> 21324-40-3

#### Cobalt Type

Component	Material	Formula / CAS
Positive Electrode	Lithium Cobalt Oxide	LiCoO <sub>2</sub> 12190-79-3
Negative Electrode	Graphite	C                7440-44-0
Electrolyte	Ethylene Carbonate – Solvent	C <sub>3</sub> H <sub>4</sub> O <sub>3</sub> 96-49-1
	Diethyl Carbonate – Solvent	C <sub>5</sub> H <sub>10</sub> O <sub>3</sub> 105-58-8
	Lithium Hexafluorophosphate – Salt	LiPF <sub>6</sub> 21324-40-3

#### Nickel Cobalt Aluminum Type

Component	Material	Formula / CAS
Positive Electrode	Lithium Cobalt Nickel Aluminum Oxide	LiCoNiAlO <sub>2</sub> 193214-24-3
Negative Electrode	Graphite	C                7440-44-0
Electrolyte	Ethylene Carbonate – Solvent	C <sub>3</sub> H <sub>4</sub> O <sub>3</sub> 96-49-1
	Diethyl Carbonate – Solvent	C <sub>5</sub> H <sub>10</sub> O <sub>3</sub> 105-58-8
	Lithium Hexafluorophosphate – Salt	LiPF <sub>6</sub> 21324-40-3

**Notice:** The information and recommendations set forth are made in good faith and are believed to be accurate at the date of preparation. Panasonic Industrial Company makes no warranty expressed or implied.



## **DISPOSAL**

All Panasonic Lithium ion batteries are classified by the federal government as non-hazardous waste and are safe for disposal in the normal municipal waste stream. These batteries, however, do contain recyclable materials. Panasonic is a Licensee of the Call2Recycle Battery Recycling Program. If you build our cells into a battery pack, please call 1-800-8-BATTERY or go to the Call2Recycle website at [www.call2recycle.org](http://www.call2recycle.org) for additional information on how your branded product can also participate in the program.

## **TRANSPORTATION**

All Panasonic lithium ion batteries are not subject to the other requirements of the US Department of Transportation (DOT) Subchapter C, Hazardous Materials Regulations if shipped in compliance with 49 CFR 173.185.

Effective January 1, 2015 all Panasonic lithium ion batteries can be shipped by air in accordance with International Civil Aviation Organization (ICAO) 2015-2016 edition, Section II or Section 1B or International Air Transport Association (IATA), 56th edition, Section II or 1B, Packing Instructions (PI) 965 (Batteries), PI 966 (Batteries, packed with equipment) and PI 967 (Batteries, contained in equipment) as appropriate.

Currently all Panasonic lithium ion batteries are regulated by the International Maritime Organization (IMO), 2012 edition, 36<sup>th</sup> amendment, under Special Provisions 188 and 230.

All Panasonic lithium ion cells are tested and comply with the UN Model Regulations, Manual of Test and Criteria, Part III, subsection 38.3.

If you build any of our lithium ion cells into a battery pack, you must also assure that they are tested in accordance with the UN Model Regulations, Manual of Test and Criteria, Part III, subsection 38.3, 5<sup>th</sup> revised edition, Amendment 1.

If you plan on transporting any untested prototype battery packs contact your Panasonic Sales Representative for regulatory information.

## **FIRST AID**

If you get electrolyte in your eyes, flush with water for 15 minutes without rubbing and immediately contact a physician. If you get electrolyte on your skin wash the area immediately with soap and water. If irritation continues, contact a physician. If the battery is ingested, call the National Capital Poison Center (NCPC) at 202-625-3333 (Collect) or your local poison center immediately.

## **GENERAL RECOMMENDATIONS**

CAUTION: Risk of fire, explosion and burns. Do not short-circuit, crush, incinerate or disassemble battery.

## **FIRE SAFETY**

In case of fire, you can use dry chemical, alcohol resistant foam or carbon dioxide fire extinguishers. Cooling the exterior of the batteries will help prevent rupturing. Burning of these batteries will generate toxic fumes. Fire fighters should use self-contained breathing apparatus. Detailed information on fighting a lithium ion battery fire can be found in Guide 147 (Lithium Ion Batteries) of the US DOT Emergency Response Guide.

PT01550



ProTechnologies  
331 ShelleyBrook Drive  
Pilot Mountain, NC 27041  
USA

## CERTIFICATE OF COMPLIANCE

**Test Data Report Number: UNLR1S1P-01**

**Characterization per the Referenced Standard: Small Lithium-Ion Battery (Single Cell)**

**Part Numbers:**

ProTechnologies	ThermoFisher	Versions Covered
PT01550	400-02692-01	All versions 7.7Wh to 11.54Wh

The design type of the product model listed has been tested in accordance with the *Recommendations on the Transport of Dangerous Goods, Manual of Tests and Criteria, 5th Revised Edition, Amendment 1, Part III, sub-section 38.3*. Test data and evaluation results are on file at ProTechnologies, Inc.

**APPROVAL DATE: 01-15-2013**

Larry Norman  
Engineering Manager

Steve Turner  
Compliance Manager